



Temperature Calibrator (CALsys 1700L) User's Guide



TEMPSENS INSTRUMENTS (I) PVT. LTD. UNIT II

A-190, Road No.5, M.I.A., Udaipur-313003 INDIA

Ph. : +91-294-3500600,

Fax : +91-294-3500631

Email : tech@tempsens.com

Web : www.tempsens.com

Legal Disclaimer

The information contained in this document is the property of TEMPSENS.

TEMPSENS reserves the right to make changes to this document and to the product described herein without notice. Before installing and using the product, review the latest version of the applicable documentation, which are available from the Tempsens website at:

<http://www.tempsens.com/>

© 2021 Tempsens Instrument Pvt. Ltd. All Rights Reserved.

Document Information

Name	Temperature Calibrator (CALsys 1700) User Manual
Document Version	1.0
Document Code	TS.CS.004
Publish Date	Saturday, December 11, 2022

Table of Contents

Preface.....	5
Safety Information.....	5
Electrical Safety.....	6
Health and Safety Instructions.....	7
Cautions and Preventions.....	8
Document Conventions.....	8
 Chapter 1	
Introduction.....	9
1.1 About CALsys 1700.....	9
1.2 Basic Working Model of CALsys 1700.....	10
1.3 Physical Measurements.....	11
1.4 Wiring Diagram.....	12
1.5 Technical Specifications for CALsys 1700L	13
1.6 Technical Specifications for CALsys 1700BB.....	14
 Chapter 2	
Setting Up CALsys 1700.....	15
2.1 Installation.....	15
2.2 Optimal Environmental Conditions.....	15
2.3 Unpacking and Initial Inspection.....	16
2.4 Operating Instructions.....	17
 Chapter 3	
Operating CALsys 1700.....	18
3.1 Turning On the Unit.....	18
3.2 Heating Up the Source.....	19
3.3 Operating Instructions.....	19
3.4 Cooling Down the Source	20

Chapter 4

Operating Unit Controller.....	21
4.1 Operation of Controller.....	21
4.2 Wiring Layout of Controller.....	22
4.3 The temperature Controller.....	23
4.4 Altering the Set Point	23
4.5 Monitoring the Controller Status	23
4.6 Temperature Units	23

Chapter 5

Digital Communication.....	24
5.1 Digital Communications Wiring.....	24
5.2 Digital Communications Parameters.....	25

Chapter 6

Software Installation.....	26
6.1 Installation.....	26
6.2 Parameters on main screen.....	26

Chapter 7

Service & Maintenance.....	28
7.1 Routine Service.....	28
7.2 Replace the Controlling Sensor.....	28
7.3 Replace Solid State Relay.....	28

Chapter 8

Troubleshooting CALsys 1700.....	30
8.1 CALsys 1700 unit does not turn on.....	30
8.2 CALsys 1700 unit is not stable.....	30
8.3 The temperature of the calibrator does not rise.....	30

Appendix A : Calibration Services	31
In House Calibration Facility.....	31
On-site Calibration Facility.....	32
Fixed-Point Calibration Facility.....	32
 Appendix B : Warranty	 33
Limit of Liability.....	33
Caution in Using the Product.....	33

Preface

Welcome to the Temperature Calibrator (CALsys 1700) user guide. This guide provides detailed information about all the product options and features, and explains how to use the product and configure basic settings to suit your requirements.

This user manual contains information about the product and its proper use and should be kept in a place where it will be easy to access. This user manual also provides safety precautions in using this product.

Safety Information

This chapter contains important information for the safety. If the instructions contained are not followed properly, particularly the safety guidelines, it could result in serious personal injury or can cause serious damage to the unit and to the components of system it is connected to. Use the instrument only as specified in this manual. Otherwise, the protection provided by the instrument may be impaired.





Refer to the safety information below and throughout the manual.

In order to assure the device operates safely, the operator needs to act effectively and be conscious of safety problems.

The terms “Warning” and “Caution” have the following definition:

- **"Warning"** indicates conditions or behaviors that could endanger the user.
- **"Caution"** denotes situations or behaviors that may endanger the instrument in use.

The following safety symbols may appear on CALsys 1700 unit:

SYMBOL	DESCRIPTION	SYMBOL	DESCRIPTION
	Risk of Danger. Important information. See Manual		Hazardous voltage. Risk of electric shock
	Caution, Hot Surface		Protective Earth Ground

Electrical Safety

WARNING:

- Before using this equipment, make sure it is properly grounded. Make sure the ground conductor wire (colored green/yellow) in the main power cable is connected to a protective earth/ground. If the equipment is not properly grounded, the high voltage may flow through the equipment body (chassis). If safety procedures are not followed, SEVERE INJURY OR DEATH may occur.
 - Do not remove the panels from the equipment without proper safety measures to avoid internal main power supply voltage hazard.
-

Follow these precautions to avoid electric shock or personal injury:

- ⊕ This equipment uses protective earth circuit to ensure that the conductive parts do not store electric charges or conduct electricity if insulation fails
- Before connecting the equipment to the electricity supply, understand the parts of the calibrator with the help of operating manual.
- Use power cables only with appropriate voltage and power rating, and that are approved for usage in your country.
- Replace the main power cable if the insulation is damaged, or if the insulation shows signs of wear and tear.
- DO NOT put the product at the location where access to the main power is blocked.
- DO NOT use an extension cord or adapter plug.
- DO NOT use the product if it operates incorrectly.
- Make sure the power cord does not touch the hot parts of the product.

Health and Safety Instructions

 **WARNING: BURN HAZARD - DO NOT touch the well access surface of the unit at high temperature**

To avoid possible health and safety concerns, follow these guidelines.

- Wear appropriate protective clothing before using the equipment.
- Operators of this equipment should be adequately trained in the handling of hot and cold items and liquids.
- Do not use the apparatus for jobs other than those for which it was designed, that is, the calibration of thermometers.
- Do not handle the apparatus when it is hot (or cold), without wearing the appropriate protective clothing and having the necessary training.
- Do not drill, modify or otherwise change the shape of the apparatus.
- Do not use the apparatus outside its recommended temperature range.
- After use, do not return the apparatus to its carrying case until the unit has cooled down.
- There are no user serviceable parts inside. When required, contact Tempsens agent for repair.
- Ensure all materials, especially flammable materials are kept away from the hot parts of the apparatus, to prevent fire risk.
- Do not use the product around explosive gas, vapor, or in damp or wet environments.
- Make sure that the space around the product meets minimum space requirements.
- DO NOT turn off the unit at temperatures higher than 350°C. This could create a hazardous situation. Select a set-point less than 100°C and allow the unit to cool before turning it off.
- The top sheet metal of the furnace may exhibit extreme temperatures for areas close to the well access.
- Materials used in this furnace may be irritating to skin, eyes, and respiratory tract. Consult the material manufacturer's Material Safety Data Sheet (MSDS) to learn about those materials before using.






Cautions and Preventions

To avoid possible damage to the instrument, follow these guidelines:

- Before working inside the equipment, turn the power off and disconnect the power cord.
- DO NOT turn the unit upside down with the inserts in place; the inserts will fall out of the unit.
- Use of this instrument at HIGH TEMPERATURES for extended periods of time requires caution.
- Completely unattended high temperature operation is not recommended for safety reasons.
- DO NOT plug the unit into 230 V if the heater switches and fuse holder reads 115 V. This action will cause the fuses to blow and may damage the instrument.
- Components and heater lifetime can be shortened by continuous high temperature operation.
- DO NOT use fluids to clean out the well.
- Never introduce foreign material into the probe hole of the insert. Fluids and other materials can leak into the calibrator causing damage.
- DO NOT drop or force the probe stems into the well. This type of action can cause a shock to the sensor and affect the calibration.

Document Conventions

The documentation uses the following conventions.

 Note:	Configuration notes
 Tip:	Recommendations or suggestions
 Important:	Information regarding required or default configuration settings and product limitations
 WARNING:	Critical actions and configuration options
 WARNING:	

Chapter 1

Introduction

1.1 About CALsys 1700

The **CALsys 1700** used as calibrator for thermocouple calibration and it can be also used as a pyrometer calibration with ceramic cavity.

The 'CALsys 1700' has been designed to provide stable and accurate temperature to enable professionals to calibrate Temperature Sensing Devices (Thermocouple & pyrometer) by comparison method up to a temperature range of 1700°C. The 'CALsys 1700L' model has been designed to be rugged and easily maintained. This model provides an isothermal enclosure (Ceramic block) in which the thermocouple can be calibrated against the temperature of the calibrator.

The 'CALsys 1700BB' model has been designed in single part. This model provides an isothermal enclosure in which the Non-contact pyrometers can be checked against the temperature of the black body. For traceable calibration a master pyrometer should be used. The units' features emissivity of 0.99, thus offering the closest approximation of a target surface that is a perfect emitter of infrared energy.



Front View



Back View


1.2 Basic Working Model of CALsys 1700

CALsys 1700' is a transportable unit designed for use on any reasonable flat surface. The target is a ridged Ceramic Tube Cavity which is painted by a high emissive paint .The target is heated by MoSi₂ heaters which allow the source to heat up to 1700°C in about 3Hrs and hold it stable at temperature within $\pm 2.0^{\circ}\text{C}$. The heater block house consists of a heater and the controller sensor. This sensor is used by the temperature controller to sense the block temperature. To maintain a required temperature the controller varies the power to the heaters via a power control device.



Important:

Before using the equipment, read the safety guidelines and operating procedures of the calibrator as described in the of this user manual.

The basic working model for CALsys 1700 is as follows:



1.3 Physical Measurement



This Page is Intestinally Left Blank

1.4 Technical Specification

CALsys 1700L has the following technical specifications:

Temperature range	500 °C to 1700 °C
Stability	±0.5°C at 500°C
	±1.0°C at 1000°C
	±1.5°C at 1700°C
Radial uniformity	±0.6°C at 500°C
	±1.4°C at 1000°C
	±1.9°C at 1700°C
Stabilization time	15 to 20mins
Controlling sensor	B type duplex
Method of Control	Self tuned PID controller
Immersion depth	185mm
Insert OD dimensions	37 mm
Heating time	3 Hrs
Resolution	1 °C
Display	LCD, °C or °F user-selectable
Size (H x W x D)	700(H) x 500(W) x 550(D) mm
Weight	Approximate 80 Kg
Power requirements	230 VAC 50/60Hz 4kW
Computer interface	RS - 232/USB
Calibration	Accredited calibration certificate provided
Environmental operating conditions	0 °C to 40 °C, 0 % to 90 % RH (non-condensing)
Specifications valid in environmental conditions	13 °C ... 33 °C

1.5 Technical Specification

CALsys 1700BB has the following technical specifications:

Temperature	500 °C to 1700 °C
Stability	±0.5°C at 500°C
	±1.0°C at 1000°C
	±1.5°C at 1700°C
Stabilization time	15 to 20 Mins
Controlling sensor	Precision B type T\C
Cavity Type	Ceramic
Cavity Dimension	29mm end closed tube
Emissivity	0.97 (± 0.02)
Heater	MoSi ₂
Heating Time	3 Hours
Method of Control	Digital self tuned PID controller
Resolution	1°C
Display	LCD, °C or °F user-selectable
Power Requirement	230 VAC 50/60Hz Range 4kW
Computer interface	RS-232/USB
Calibration	Accredited calibration certificate provided
Dimensions	700(H) x 500(W) x 550(D) mm
Weight	Approximate 80 Kg

Chapter 2

Setting up CALsys 1700

2.1 Installation

Place the calibrator on a flat surface with at least 10 inches of free space around the instrument. Overhead clearance is required.

DO NOT Place this unit under a cabinet or structure. Plug the power cord into a grounded mains outlet located on the controlling unit rear panel. Observe that the nominal voltage corresponds to that indicated in the [Technical Specifications](#) in [Chapter 1](#) of this user's guide.

2.2 Optimal Environmental Conditions

Although the instrument has been designed for optimum durability and trouble-free operation, it must be handled with care. The instrument should not be operated in an excessively dusty or dirty environment. Refer to [Chapter 8, Service & Maintenance](#) in this user's manual for routine service and cleaning recommendations.

The instrument operates safely under the following conditions:

- Temperature range: 5°C - 50°C (41°F - 122°F)
- Ambient relative humidity: 15 - 50%
- Pressure: 75kPa - 106kPa
- Mains voltage within $\pm 10\%$ of nominal
- Vibrations in the calibration environment should be minimized
- Altitude less than 2000 meters

2.3 Unpacking and Initial Inspection

CALsys 1700 is packed in custom-designed packaging to send out your unit. Unpack the furnace carefully. Inspect the unit after unpacking for any signs of damage, and confirm that your delivery is in accordance with the packing note. If you find any damage to the unit or an item is missing, notify Tempsens immediately.

The following accessories are included in the package:

- CALsys 1700L/BB
- Heating Chamber
- Heaters (MoSi₂)
- RS-232 Cable
- B type thermocouple
- Ceramic Cavity/optional block for dry block
- Manual
- Certificates
- Insulation Wool



CALsys 1700L/BB



Heating Chamber



Heaters




Insulation Wool



Cavity



RS-232



B Type Thermocouple



Optional Block

2.4 Operating Instructions

1. Please open the wooden box carefully and takeout the operating manual from the box and read carefully.
2. Take out the temperature calibrator unit carefully and keep it at suitable place.
3. Install inner chamber as instructed in [Installation of CALsys 1700](#) in [Chapter -3](#) of this user guide.
4. Connect the power plug to the mains.
5. Turn the switch ON, and observe the display on the controller. The display shows that the calibrator is ready for use.
6. Keep the switch in the ON position
7. Ensure the ceramic insert/ceramic cavity is properly inserted in the calibrator.



Ceramic Cavity for 1700BB



Ceramic Block for 1700L



Note:


- Ceramic insert or ceramic cavity should be removed from the calibrator when it will reach ambient temperature.
 - Use either block or ceramic cavity (DO NOT use them together).
 - Always use safety gloves while handling the calibrator.
-

Chapter 3

Operating CALsys 1700


3.1 Turning ON the unit

Plug the CALsys 1700L power cord into mains outlet of the proper voltage, frequency and current capability. Typically this will be (230 VAC \pm 10, 50/60 Hz). Using the “ON/OFF” (for heater supply) switch located at front side responsible for heating of furnace. The furnace will turn on and begin to heat the previously programmed temperature set-point.



3.2 Heating up the source

Press “UP” or “DOWN” key of controller to change the set-point value. When the set-point temperature is changed the controller will switch the furnace heater on or off to raise or lower the temperature. The displayed temperature will gradually change until it reaches the set-point temperature. The furnace may require time to reach the set-point depending on the span. An another 5 to 10 minutes is required to stabilize the furnace with ± 1.5 Deg C of the set-point.



! Important:

- All other controller parameters are company set and locked. It is recommended not to change them.
- When the source is operated at any temperature above ambient, the front face and plate become hot.
- Always put the fan in “ON” Condition and do not switch the “Mains” off Directly, First set the controller to 0°C and then wait until the unit is not cooled below 350°C.

3.3 Operating Instruction

1. Connect the 'CALsys 1700BB' to a suitable power supply & set the controller to the required Temperature.
2. After stabilization of the furnace, aim the reference standard thermometer to the center of cavity and take the readings.
3. Repeat step 2 with the thermometer which is under calibration.
4. Find out the error by comparison method.
5. Always use a reference IR thermometer for comparison calibration method for 1700BB.
6. Check the D: S Ratio of the Pyrometer, it must be $>10:1$.
7. For CALsys 1700L place the ceramic insert in the calibrator.
8. Place the sensors for calibration and master sensor into a suitable insert hole.

9. If test and master sensor are thermocouple, then always use compensating cable (for each type such as J, K, T, E, N, R, S, B) for interconnection between sensor and Digital Multimeter (DMM).
10. Keep reference junction at 0°C. If keeping at 0°C is not possible, then add millivolt (mV) of room (ambient) temperature with sensors output (mV) to compensate the ambient temperature.
11. When controller temperature becomes stable, record the reading from DMM with at least 4 1/2 digit precision, which should be calibrated for the measuring range (mV & ohms measurements).
12. Find out the error by comparison method.
13. Reset the controller and/or repeat the calibration for another calibration point or for another sensor.
14. Reset the controller and / or repeat the calibration for another calibration point or for another sensor.
15. When the calibration is completed, reset the controller to 0°C & wait until the unit to cool below 350°C, before moving the 'CALsys 1700L/BB' to new location.



Note:

- Ceramic Insert should be of required hole size so that the hole size is suitable for sensor under calibration.
- Check that all sensors (test and master) are immersed at same depth in insert (ceramic block).
- Plan the calibration point by gradually increasing temperature, then set the controller accordingly. This method will save the time.
- Take the reading of sensors (master and test) at stable temperature with the help of digital multi-meter, or by using Tempsens' TEMPMET 08 or TEMPMET 09 for reading RTD or TC sensor.

3.4 Cooling Down the Source

- Before transporting the ceramic block, ensure that the temperature of black body has cooled sufficiently.
- If you require black body to cool quickly, set the temperature to the room temperature (ambient temperature).
- Always cool the heated instrument to ambient temperature before disconnecting it from the mains, switching it OFF, or removing the temperature sensor or test item.
- Always keep the heated instrument supervised and under observation until it cools down sufficiently.

Chapter 4

Operating Unit Controller

4.1 Operation Of controller

Plug the CALsys 1700L power cord into mains outlet of the proper voltage, frequency and current capability. Typically this will be (230 VAC±10, 50/60 Hz). Using the “ON/OFF” (for heater supply) switch located at front side responsible for heating of furnace. The furnace will turn on and begin to heat the previously programmed temperature set-point.











P104 Controller

Beacons 1234		Operator Buttons	
ALM	Alarm active (Red)		Scroll button. Press to scroll forward through a list of parameters. Hold down to scroll continuously.
1	Lit when output 1 is ON (typically heating)		Page button. Press to scroll back through a list of parameters. Hold down to select a different operating level. This button can be assigned a specific function - see 'P' code P73 section 5.2.18.
2	Lit when output 2 is ON (typically cooling)		Press to decrease a value.
3	Lit when output 3 is ON (P108 & P104 only)		Press to increase a value.
4	Lit when output 4 is ON (typically alarm)	Press and together to return to the operating display.	
SPX	Alternative setpoint in use (SP2)	F1	Function key 1
REM	Digital communications active (flashing)	F2	Function key 2
MAN	Manual mode selected	These buttons are only available in P108 & P104. For functionality see section 5.2.18	

Alpha-Numeric Display		
Top row	Measured Temperature (Process Value, PV) or the value of a selected parameter	
* Second Line	Target Temperature (Setpoint, SP) or the mnemonic of a selected parameter	These are the default parameters. They may be customised to show alternative parameters to suit the requirements of a particular process, see section 5.2.19.
* Third Line	Output power demand	

4.2 Wiring Layout of Controller



Key to symbols used in wiring diagrams					
	Logic (SSR drive) output		Relay output		Changeover relay output
	0-20 or 4-20mA analogue output isolated		Triac output		
	Current transformer input		Contact input		

4.3 The Temperature Controller

The upper display of the controller indicates the measured temperature, the middle display indicates the desired temperature or set point and lower display shows output power demand.

4.4 Altering the Set point

To change the set point of the controller simply use the UP and DOWN keys to raise and lower the set point to the required value. The lower display changes to indicate the new set point.

4.5 Monitoring the Controller Status

A row of beacons indicate the controllers status as follows,

OP1 Heat Output

OP2 Cool Output (only for models which operate below 0°C)

REM This beacon indicates activity on the PC interface

4.6 Temperature Units

Momentary pressing of the Scroll key will show the controller units °C or °F.



IMPORTANT NOTICE

The controller's function settings are preset and will not require adjustment.

Chapter 5

Digital Communication


Digital Communication allows the controller to communicate with a PC or a networked computer system through RS-232 (or EIA232) protocol. RS-232 (or EIA232) is a standard communication protocol for linking computer and its peripheral devices to allow serial data exchange. RS-232 communication is not available if Remote Set point is fitted. The cable screen should be grounded at one point only to prevent earth loops.

5.1 Digital Communications Wiring

To use EIA232 (or RS-232), the PC will be equipped with an EIA232 port, usually referred to as COM 1. To construct a cable for EIA232 operation use a three core screened cable. The terminals used for EIA232 digital communications are listed in the table below. Some PC's use a 25-way connector although the 9 way is more common.

STANDARD CABLE COLOR	PC SOCKET PIN NO.		PC FUNCTION*	INSTRUMENT TERMINAL	INSTRUMENT FUNCTION
	9 WAY	25 WAY			
White	2	3	Receive (RX)	HF	Transmit (TX)
Black	3	2	Transmit (TX)	HE	Receive (RX)
Red	5	7	Common	HD	Common
Link Together	1	6	Received line signal detect		
	4	8	Data terminal ready		
	6	11	Data set ready		
Link Together	7	4	Request to send		
	8	5	Clear to send		
Screen		1	Ground		

*These are the functions that are normally assigned to socket pins. Check your PC manual to confirm.



5.2 Digital Communications Parameters

The following table shows the available parameters.

DIGITAL COMMUNICATION LIST (COMMS)						
NAME	SCROLLING DISPLAY	PARAMETER DESCRIPTION	VALUE		DEFAULT	ACCESS LEVEL
ID	MODULE IDENTITY	COMMUNICATION IDENTITY	none	NO MODULE FITTED	AS ORDER CODE	CONF L3 R/O
			R232	RS232 MODBUS INTERFACE		
			R485	EIA485 MODBUS INTERFACE		
			R422	EIA422 MODBUS 3216 ONLY		
			DCP	REMOTE SET-POINT INPUT. IF FITTED THIS ID REPLACES THE ABOVE AND NO FURTHER PARAMETERS ARE SHOWN		
ADDR	COMMUNICATION ADDRESS	COMMUNICATION ADDRESS OF THE INSTRUMENT	1 TO 254		1	L3
BAUD	COMMUNICATION BAUD RATE	COMMUNICATION BAUD RATE	1200	1200	9600	
			2400	2400		
			4800	4800		
			9600	9600		
			19.20	19.20		
PRTY	COMMUNICATION PARITY	COMMUNICATION PARITY	none	NO PARITY	NONE	
			Even	EVEN PARITY		
			Odd	ODD PARITY		
DELAY	RX/TX DELAY TIME	TO INSERT A DELAY BETWEEN RECEIVER (Rx) AND TRANSMITTER (Tx) TO ENSURE THAT DRIVERS HAVE SUFFICIENT TIME TO SWITCH OVER.	OFF	NO DELAY		CONF L3 R/O
			On	FIXED DELAY APPLIED		
RE TRAN	COMMS RETRANSMISSION	MASTER COMMS BROADCAST PARAMETER SEE SECTION 15.2.1	none	NONE	NONE	
			WSP	WORKING SETPOINT		
			PU	PROCESS VARIABLE		
			OP	OUTPUT DEMAND		
			ERR	ERROR		
REGAD	COMMS RETRANSMISSION ADDRESS	PARAMETER ADDED IN SLAVE ADDRESS TO WHICH THE MASTER COMMUNICATION VALUE WILL BE WRITTEN	0 TO 9999			

Chapter 6

Software Installation

The provided Tempsens software offers possibilities to connect furnace temperature bath and change set point, maximum time span, view real time graph and evaluate measuring data.

6.1 Installation

Install the calibration software using the installation guide file. After installing the software, start the application. CalSoft Temsense Instruments screen appears, as shown in .



Figure 1 - Main screen



Figure 2 - PC ComPort



Figure 3 - Controller Version

6.2 Parameters on Main Screen

Communication between the furnace and the software is implemented via RS-232 cable connected between the furnace and the PC serial port. This enables the acquisition and recording of data, as well as the transfer of commands from the software application to the Tempsens furnace.

To communicate between the furnace and software application, select correct **COM port** (as shown in *Figure 2*) and Controller Version (**2604**) (as shown in *Figure 3*), and then click **Connect**.

To set the temperature of furnace as per your requirement, type the required temperature in **Set Point (°C)** field, and click **Set**.

PVI Value displays the current present value of furnace temperature.

The **Scale Trend** section enables you to configure **Y-Axis Min** (minimum value: **0**), **Y-Axis Max** (maximum value: **1500**), **Maximum Time Span** of data logging up to 120 minutes, and then click the **Start Graph** button. After completing the task, click **Save To File** to save settings to file. See *Figure 4*.



Figure 4 - Scale Trend Window

File will be stored in **.xls** format to save previous record. See *Figure 5*.




Figure 5 - Record in .xls Format

Chapter 7

Service & Maintenance

7.1 Routine Service

Turn the electricity supply off before attempting any cleaning operation. The only moving part is the fan. That has sealed-for-life bearings. Depending on the environment in which it is used, periodic cleaning is recommended. Cleaning may be accomplished by the use of a small dry paint brush. The instrument should be periodically checked to ensure it is in good order both mechanically and electrically.

7.2 Replace the Controlling Sensor

Replacement of the controlling T/C Sensor is carried out as follows:

1. Make sure that the Calibrator Unit is not connected to the mains supply.
2. Remove the front and back cover of the Calibrator. Refer Page 20.
3. Disconnect the T/C from the M & S Terminal. Refer Page 28.
4. Remove the T/C from the heating chamber present inside by unscrewing the Screw fitted in the fitting. . Refer Page 26.
5. Fit the new T/C.
6. Reconnect the new T/C to the connector block (M & S Plate)..
7. Replace the cover.

7.3 Replace Solid State Relay

1. Disconnect the Calibrator Unit from the electrical supply.
2. Remove the cover of the furnace.
3. Locate the solid state relay.
4. Disconnect the wires, noting their position.
5. Remove the faulty relay and replace it with a new one.
6. Tighten the fixing screws.
7. Refit the wires has been noted in step 4.
8. Replace the cover.
9. Reconnect the furnace to the electrical supply.



Note:

- The calibration instrument has been designed with the utmost care. Ease of operation and simplicity of maintenance have been a central theme in the product development. Therefore, with proper care the instrument should require very little maintenance. Avoid operating the instrument in an oily, wet, dirty, or dusty environment.
 - If the outside of the instrument becomes soiled, it may be wiped clean with a damp cloth and mild detergent. Do not use harsh chemicals on the surface which may damage the paint.
 - It is important to keep the calibrator clean and clear of any foreign matter. DO NOT use chemicals to clean the calibrator.
 - Avoid knocking or dropping the instrument.
 - If a hazardous material is spilt on or inside the equipment, the user is responsible for taking the appropriate decontamination steps as outlined by the national safety council with respect to the material.
 - If the mains supply cord becomes damaged, replace it with a cord with the appropriate gauge wire for the current of the instrument. If there are any questions, contact tempsens.
-

Chapter 8

Troubleshooting CALsys 1700

8.1 CALsys 1700 unit does not turn on

If the CALsys 1700 unit does not turn on or operate as usual, check if the fuse is broken, and replace the fuse if necessary. If the fuse breaks repeatedly, contact Tempsens for technical support.

8.2 The CALsys 1700 unit is not stable

If the CALsys 1700 is not stable or the temperature varies too often, the control parameters might have been modified. If this happens, contact Tempsens for technical support.

8.3 The temperature of the calibrator does not rise

If the temperature of the Calibrator Unit does not rise, ensure the following:

- (a) The power cord is tightly fitted at the power entry location on the Calibrator Unit.
- (b) The switch should be in ON position.

If the temperature of the Calibrator Unit still does not rise even after you have inspected and fitted the power cord and power switch, the reason can be one of the following:

- The heating element is not firmly closed. If the heating element is open, close it properly.
- The thermocouple is not firmly closed. If the thermocouple is open, close it properly.
- The Controller Unit may be not display correct reading.
- The SSR is damaged.

If the Calibrator Unit does not function properly, contact Tempsens immediately for technical support.

Appendix A: Calibration Services

Tempsens Calibration Center is an independent unit of Tempsens instruments (I) Pvt. Ltd, having laboratories at Udaipur, Vadodara & Bangalore. It is accredited for wide range of temperature calibration services.

It is the only private sector Laboratory in the country with accredited Fixed Point Temperature Calibration Facilities. The lab has highly stable calibration furnaces, measuring instruments and accurate master sensors traceable to National and International Standards.

The calibration center functions as per ISO 17025 / NABL standards. Calibration of contact type sensors can be made in temperature range of -196°C to 1600°C and Calibration of non-contact type sensors can be made in temperature range 0°C to 2900°C. Further the laboratory is accredited for onsite temperature calibration.

The lab offers both at Lab & On-Site Calibration of Furnace/Bath from -80°C to 1600°C and Black Body Calibration from 50°C to 1700°C. Furnace/Chamber Calibration (TUS) with multiple sensors from -80°C to 1200°C is also in the scope of the lab.

In House Calibration Facility

QUALITY MEASURED / INSTRUMENTS	TEMPERATURE RANGE	CALIBRATION & MEASUREMENT CAPABILITY
Contact Type RTD, Thermocouples Thermometers	-196°C	0.05°C
	-80 to -38°C	0.03°C
	-38°C to 0°C	0.03°C
	>0°C to 140°C	0.04°C
	>140°C to 250°C	0.04°C
	>250°C to 650°C	0.12°C
	>650°C to 1200°C	1.30°C
	>1200°C to 1600°C	2.60°C
Non-Contact Type Pyrometer	0°C to 250°C	1.5°C
	>250°C to 500°C	2.4°C
	>500°C to 1500°C	2.5°C
	>1500°C to 1700°C	3.2°C
	>1700°C to 2900°C	4.0°C

On-site Calibration Facility

QUALITY MEASURED / INSTRUMENTS	TEMPERATURE RANGE	CALIBRATION & MEASUREMENT CAPABILITY
Contact Type RTD, Thermocouples Thermometers	-25°C to 0°C >0°C to 140°C >140°C to 250°C >250°C to 650°C >650°C to 1200°C	0.07°C 0.04°C 0.09°C 0.12°C 1.30°C
Non-Contact Type Pyrometer	0°C to 250°C >250°C to 500°C >500°C to 1200°C	1.50°C 2.40°C 2.5°C
Multipoint Position Calibration of Chamber, Oven, Furnaces (Thermal Mapping(TUS))	-80°C to 200°C >200°C to 1200°C	2.8°C 4.1°C

Fixed-point Calibration Facilities

QUALITY MEASURED / INSTRUMENTS	TEMPERATURE RANGE	CALIBRATION & MEASUREMENT CAPABILITY
Calibration of SPRT/PRTS/thermocouple and so on.	Triple Point of Water (0.01°C) Melting Point of Gallium (29.7646°C) Freezing Point of Tin (231.928°C) Freezing Point of Zinc (419.527°C) Freezing Point of Aluminum (660.323°C)	0.0038°C 0.0065°C 0.0065°C 0.0071°C 0.0075°C
Calibration of Thermocouple at Secondary Fixed Point	Melting Point of Gold (1064.18 °C)>1500°C to 1700°C >1700°C to 2900°C	0.72°C2.5°C 3.2°C 4.0°C
	Melting Point of Palladium(1554.8	0.83°C

Appendix B: Warranty

This instrument has been manufactured to exacting standards and is warranted for twelve months against electrical breakdown or mechanical failure caused through defective material or workmanship, provided the failure is not the result of misuse. In the event of failure covered by this warranty, the instrument must be returned, carriage paid, to the supplier for examination and will be replaced or repaired at our option.

FRAGILE CERAMIC AND/OR GLASS PARTS ARE NOT COVERED BY THIS GUARANTEE

INTERFERENCE WITH OR FAILURE TO PROPERLY MAINTAIN THIS INSTRUMENT MAY INVALIDATE THIS GUARANTEE

Limit of Liability

TEMPSENS is not liable for any damages that arise from the use of any examples or processes mentioned in these Specifications are subject to change without notice.

Caution in Using the Product

TEMPSENS PRODUCTS ARE INTENDED FOR USE BY TECHNICALLY TRAINED AND COMPETENT PERSONNEL FAMILIAR WITH GOOD MEASUREMENT PRACTICES.

IT IS EXPECTED THAT PERSONNEL USING THIS EQUIPMENT WILL BE COMPETENT WITH THE MANAGEMENT OF APPARATUS WHICH MAY BE POWERED OR UNDER EXTREMES OF TEMPERATURE, AND ARE ABLE TO APPRECIATE THE HAZARDS WHICH MAY BE ASSOCIATED WITH, AND THE PRECAUTIONS TO BE TAKEN WITH, SUCH EQUIPMENT

Note:

[illegible]



www.tempsens.com

TEMPSENS INSTRUMENTS (I) PVT. LTD.

Tempsens Instruments (I) Pvt. Ltd. U# I
B-188A, Road No.5, M.I.A., Udaipur-313003 (Rajsthan) INDIA

Ph. : +91-294-3507700, **Fax :** +91-294-3507731

Email : info@tempsens.com